ACCU-CHEK® Instant S

Roche

USER'S MANUAL

BLOOD GLUCOSE METER

This file may not print or view at 100%.
Die lines and color breaks do not print.
"Holding Area" text, box and rules do not print.

Roche USA – 83505
V1/1 – 07945914061A
07945914061(01) – Black

PMS-MANUAL FOLD OUT

Approved according to SHR_SOP 04.07.08_Att A **Category 2**

The Signature will not be printed!

LAN	holeckos - Aug 10, 2021
IPM	emrichs - Aug 24, 2021
PM/LAB	smithk44 - Aug 13, 2021
GD/LAY	brownind - Aug 24, 2021

Contents

Introduction 2

Chapter 1: Your New System 5

Chapter 2: Blood Glucose Tests 9

Chapter 3: Control Tests 15

Chapter 4: Review Your Data 19

Chapter 5: Connecting to a PC 21

Chapter 6: Meter Maintenance and Troubleshooting 22

Chapter 7: Technical Information 29

Chapter 8: Guarantee 34

Index 35

5_07945914061A_07945914061_01.indd 1 8/4/2021 11:04:46 AM

Introduction

This User's Manual contains warnings, precautions and notes:

⚠ WARNING

A **WARNING** indicates a foreseeable serious hazard.

⚠ PRECAUTION

A **PRECAUTION** describes a measure you should take to use the product safely and effectively or to prevent damage to the product.

NOTE

A **NOTE** draws your attention to important information to help you get the most out of using the product.

The Accu-Chek Instant S System

The Accu-Chek Instant S meter with the Accu-Chek Instant test strips is indicated to quantitatively measure glucose in fresh capillary whole blood from the finger, palm, forearm and upper arm as an aid in monitoring the effectiveness of glucose control.

The Accu-Chek Instant S meter with the Accu-Chek Instant test strips is intended for in vitro diagnostic self-testing by people with diabetes.

The Accu-Chek Instant S meter with the Accu-Chek Instant test strips is intended for in vitro diagnostic near-patient testing by healthcare professionals in clinical settings. Venous, arterial and neonatal blood testing is limited to healthcare professional use.

This system is not for use in diagnosis of diabetes mellitus, nor for testing neonate cord blood samples.

The system includes:

Accu-Chek Instant S meter with battery, Accu-Chek Instant test strips* and Accu-Chek Instant control solutions*.

*Some items may not be included in the kit. They are a separate purchase.

⚠ WARNING – KEEP BATTERIES OUT OF REACH OF CHILDREN

In Australia: If you suspect your child has swallowed or inserted a button battery immediately call the Australian 24-hour Poisons Information Centre on 13 11 26 for fast, expert advice.

5_07945914061A_07945914061_01.indd 2 8/4/2021 11:04:46 AM

⚠ WARNING

Risk of suffocation
This product contains small parts that can be swallowed. Keep the small parts away from small children and people who might swallow small parts.

Risk of life-threatening injuries

- Keep new and used batteries away from children. Ingestion or insertion into the body may cause chemical burns, perforation of soft tissues and death. Severe burns may occur within 2 hours of swallowing. If you think a battery might have been swallowed or placed inside any part of the body, seek medical attention immediately.
- If the battery compartment does not close securely, stop using the product and keep it away from children. Contact Roche.

⚠ WARNING

Risk of infection
Any object coming into contact with human blood is a potential source of infection (see: Clinical and Laboratory Standards Institute: Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline – Fourth Edition; CLSI document M29-A4, May 2014). Healthcare Professionals: see also the Information for Healthcare Professionals section in the chapter Technical Information.

Risk of a serious health incident
Failure to follow testing instructions or test strip storage and handling instructions can lead to an incorrect test result that may lead to improper therapy. Carefully read and follow the instructions in the User's Manual and package inserts for the test strips and control solutions.

5_07945914061A_07945914061_01.indd 3 8/4/2021 11:04:48 AM

Introduction

Important Information About Your New Meter

- It is strongly recommended to have a back-up testing method available. Failure to test could cause a delay in therapy decisions and lead to a serious medical condition. Examples of back-up testing methods include a back-up meter and test strips. Ask your healthcare professional or pharmacist about other possible back-up methods.
- Sample data screens are shown throughout the manual. Your data will differ.

Reporting of Serious Incidents

For a patient/user/third party in the European Union and in countries with identical regulatory regime; if, during the use of this device or as a result of its use, a serious incident has occurred, please report it to the manufacturer and to your national authority.

5_07945914061A_07945914061_01.indd 4 8/4/2021 11:04:48 AM

1 Your New System

The Accu-Chek Instant S Meter



Right Side View | Front View | Left Side View

1. **Meter Button**
   Press to turn the meter on and off, move to the next screen or make a selection.

2. **Display**
   Shows results, messages and test results stored in memory.

3. **Target Range Indicator**
   Indicates whether your blood glucose result falls above, within or below a pre-set range of values.

4. **Micro USB Port**
   Transfers data from the meter to a computer (PC).

5_07945914061A_07945914061_01.indd 5 8/4/2021 11:04:52 AM

1 Your New System

Back View

Underside View

5. Battery Door
Open to replace battery.

6. Test Strip Slot
Insert test strip here.

7. Test Strip Container*
(for example)

8. Test Strip - Metallic End
Insert this end into the meter.

9. Test Strip - Yellow Edge
Touch blood drop or control solution here.

10. Control Solution Bottle*

11. Battery

*Some items may not be included in the kit. They are a separate purchase.

5_07945914061A_07945914061_01.indd 6 8/4/2021 11:04:58 AM

Using the Meter Button

Your new meter has only one button, but it performs many actions. Press the meter button once to turn the meter on and see your most recent blood glucose result. Continue pressing to view your 7, 30 and 90-day blood glucose averages.

When instructed to press the meter button, press it briefly and release it.

When instructed to press and **hold** the meter button, press and hold it for 3 or more seconds.

Time and Date

To display or change the time and date on the meter, connect it to a PC with diabetes management software installed. Follow the diabetes management software instructions to set the time and date. The meter display will then show the time and date along with your test results.

Target Range

NOTE

The target range indicator is not a substitute for hypoglycaemia training by your healthcare professional.

Consult your healthcare professional for the blood glucose range appropriate for you. It is very important to stay within your target range.

The target range is set by default at 70–180 mg/dL (3.9–10.0 mmol/L). The target range can be set from a lower limit of 60–140 mg/dL (3.3–7.8 mmol/L) to an upper limit of 90–300 mg/dL (5.0–16.6 mmol/L).

To change the target range, connect the meter to a PC with diabetes management software installed. Then follow the diabetes management software instructions.

5_07945914061A_07945914061_01.indd 7 8/4/2021 11:04:58 AM

1 Your New System

Symbols

Symbol	Description
	Last test result symbol. This indicates that you are viewing your last blood glucose result.
7 day ave	7-day average symbol. This indicates that you are viewing the 7-day average of your blood glucose results.
30 day ave	30-day average symbol. This indicates that you are viewing the 30-day average of your blood glucose results.
90 day ave	90-day average symbol. This indicates that you are viewing the 90-day average of your blood glucose results.
C	Control bottle symbol
	Drop symbol
	Hourglass symbol
	Low battery symbol
- - -	No data to display
	Target range indicator arrow. When flashing, your blood glucose result is outside the target range.
	Temperature warning symbol

5_07945914061A_07945914061_01.indd 8 8/4/2021 11:05:15 AM

⚠ WARNING

Risk of a serious health incident
Using the wrong unit of measurement may cause misinterpretation of your actual blood glucose level and may lead to improper therapy. Blood glucose results can be displayed in either mg/dL or mmol/L. The back label of the meter shows the unit of measurement. If the meter shows the wrong unit, contact Roche. If you do not know which unit of measurement is correct for you, contact your healthcare professional.

mg/dL or mmol/L is printed here

Using the Accu-Chek Instant S System

⚠ WARNING

Risk of a serious health incident

- If you drop the meter or drop the meter with a test strip inserted, the meter and/or test strip could be damaged. Discard the test strip and perform a control test with control solution and a new, unused test strip to ensure the meter and test strips are working properly. Then repeat the blood glucose test with a new test strip.
- A test strip that is not stored or used properly can deliver an incorrect test result.
- Do not store test strips in high heat and moisture areas (bathroom or kitchen)! Heat and moisture can damage test strips.

- Use only Accu-Chek Instant test strips.
- Use the test strip immediately after removing it from the test strip container.
- Do not apply blood or control solution to the test strip before inserting it into the meter.
- Close the test strip container tightly immediately after removing a test strip to protect the test strips from humidity.
- Store the unused test strips in their original test strip container with the cap closed.

5_07945914061A_07945914061_01.indd 9 8/4/2021 11:05:16 AM

2 Blood Glucose Tests

- Check the use by date on the test strip container. Do not use the test strips after that date.
- Store the test strip container and meter in a cool, dry place such as a bedroom.
- Refer to the test strip package insert for test strip storage and system operating conditions.

Performing a Blood Glucose Test with Blood from Your Fingertip

NOTE

- You need the meter, a test strip and a lancing device with a lancet inserted to perform a blood glucose test.
- A blood glucose test cannot be performed while the meter is connected to a PC with a USB cable.

1

Wash your hands with warm soapy water and dry thoroughly.

Prepare the lancing device.

2

Check the use by date on the test strip container.

Do not use test strips past the use by date.

3

Remove a test strip from the test strip container.

Close the cap tightly.

4

Insert the metallic end of the test strip into the meter. The meter turns on.

5



When a flashing drop symbol appears, perform a fingerstick with the lancing device.

6

Gently squeeze your finger to assist the blood flow. This helps you get a blood drop.

5_07945914061A_07945914061_01.indd 10 8/4/2021 11:05:22 AM

7

Touch the **yellow edge** of the test strip to the blood drop.

Do not put blood on top of the test strip.

8

Remove your finger from the test strip when the flashing hourglass symbol appears.

9

or



Your test result appears with an arrow. This arrow shows if your test result falls above, within or below the target range.* The target range is represented by the green region of the target range indicator. The arrow will flash if your test result falls above or below this range.

Remove and discard the used test strip.

*The target range is set by default at 70–180 mg/dL (3.9–10.0 mmol/L). To change the target range, connect the meter to a PC with diabetes management software installed. Consult your healthcare professional before changing the target range.

5_07945914061A_07945914061_01.indd 11 8/4/2021 11:05:30 AM

NOTE

When performing a **blood glucose test**: If the control bottle symbol and the flashing L1 or L2 appear on the display with your blood glucose result, an error has occurred.

- Do not act on the blood glucose result.
- Discard the test strip and repeat the blood glucose test with a new test strip.

Performing a Blood Glucose Test with Blood from Your Palm, Forearm or Upper Arm (Alternative Site Testing)

⚠ WARNING

Risk of a serious health incident
Your blood glucose level changes faster in your fingertip and palm than in the forearm and upper arm. Performing a blood glucose test with blood from the forearm or upper arm may cause you to misinterpret your actual blood glucose level, leading to improper therapy.

- Do not use alternative site testing (AST) to calibrate a continuous glucose monitoring system.
- Do not use alternative site testing to make insulin dosing calculations.
- Alternative site testing should be done only during steady-state times (when glucose is not changing rapidly).

You have the option of obtaining a blood sample from other sites on your body besides the fingertip. Alternative sites include the palm, forearm and upper arm.

Blood obtained from the fingertip and palm can be used at any time to perform a blood glucose test.

If blood from the forearm or upper arm is used, there are certain times when testing is not appropriate. Read the following section before you try testing from the forearm or upper arm.

You may perform a forearm or upper arm test	• immediately before a meal. • while fasting.

5_07945914061A_07945914061_01.indd 12 8/4/2021 11:05:30 AM

You may NOT perform a forearm or upper arm test	• up to 2 hours following a meal, when blood glucose values can rise quickly. • after injecting bolus insulin, when blood glucose values can decrease rapidly. • after exercise. • if you are sick. • if you think your blood glucose is low (hypoglycaemia). • if you sometimes do not notice when your blood glucose is low.

If you are interested in AST, talk to your healthcare professional first.

To obtain an AST cap and detailed AST instructions, contact Roche.

Unusual Blood Glucose Results

⚠ PRECAUTION

Risk of a serious health incident
If your blood glucose results do not match how you feel, check the following list to help solve the problem.

Troubleshooting Checks	Action
Did you wash your hands?	Wash your hands with warm soapy water and dry thoroughly. Repeat the blood glucose test with a new test strip.
Were the test strips expired?	Discard the test strips if they are past the use by date. Repeat the blood glucose test with an unexpired test strip.
Was the cap on the test strip container always closed tightly?	Replace the test strips if you think the test strip container was uncapped for some time. Repeat the blood glucose test.
Was the test strip used immediately after it was removed from the test strip container?	Repeat the blood glucose test with a new test strip.
Were the test strips stored in a cool, dry place?	Repeat the blood glucose test with a properly stored test strip.

5_07945914061A_07945914061_01.indd 13 8/4/2021 11:05:30 AM

Troubleshooting Checks	Action
Did you follow the directions?	Read the chapter Blood Glucose Tests and repeat the blood glucose test. Contact Roche if you still have problems.
Are the meter and test strips working properly?	Perform a control test. See the chapter Control Tests for instructions.
Are you still unsure of the problem?	Contact Roche.

Symptoms of Low or High Blood Glucose

⚠ WARNING

Risk of hypoglycaemia
Being aware of the symptoms of low or high blood glucose can help you understand your test results and decide what to do if they seem unusual.

- If you are experiencing any of the following symptoms or other unusual symptoms, test your blood glucose from the fingertip or palm.
- If your blood glucose result is displayed as LO or HI, contact your healthcare professional immediately.

Low blood glucose (hypoglycaemia): Symptoms of hypoglycaemia may include, but are not limited to, anxiety, shakiness, sweating, headache, increased hunger, dizziness, pale skin colour, sudden change in mood or irritability, fatigue, difficulty concentrating, clumsiness, palpitations and/or confusion.

High blood glucose (hyperglycaemia): Symptoms of hyperglycaemia may include, but are not limited to, increased thirst, frequent urination, blurred vision, drowsiness and/or unexplained weight loss.

Disease Effects and Prevalence

For information on the effects and prevalence of diabetes in your area, visit the International Diabetes Federation website at www.idf.org or send an email to info@idf.org. For further advice or helpline information, refer to the national diabetes organisation for your country.

5_07945914061A_07945914061_01.indd 14 8/4/2021 11:05:30 AM

3 Control Tests

When to Perform a Control Test

Performing a control test lets you know the meter and test strips are working properly. You should perform a control test when:

- you open a new test strip box.
- you left the test strip container open.
- you think the test strips are damaged.
- you want to check the meter and test strips.
- the test strips were stored in extreme temperatures, humidity or both.
- you dropped the meter.
- your test result does not match how you feel.
- you want to check if you are performing the test correctly.

About the Control Solutions

- Use only Accu-Chek Instant control solutions.
- Close the control solution bottle tightly after use.
- Write the date you open the control solution bottle on the bottle label. The control solution must be discarded 3 months from the date the control solution bottle was opened (discard date) or on the use by date on the bottle label, whichever comes first.
- Do not use control solution that is past the use by or discard date.
- Refer to the control solution package insert for control solution storage conditions.
- The meter automatically recognises the difference between the control solution and blood.
- The control solution can stain fabric. Remove stains by washing with soap and water.

Performing a Control Test

You need the meter, a test strip and control solution Level 1 or Level 2.

1

Check the use by date on the test strip container. Do not use test strips past the use by date.

2

Remove a test strip from the test strip container.

Close the cap tightly.

3



Insert the metallic end of the test strip into the meter. The meter turns on.

A flashing drop symbol appears.

5_07945914061A_07945914061_01.indd 15 8/4/2021 11:05:32 AM

4

Select the control solution to test. You will enter the level later in the test.

5

Remove the bottle cap. Wipe the tip of the bottle with a tissue. Squeeze the bottle until a tiny drop forms at the tip.

6



Touch the drop to the **yellow edge** of the test strip. Do not put control solution on top of the test strip.

Remove the control bottle from the test strip when the flashing hourglass symbol appears.

7

or

The control result appears. Press the meter button.

NOTE

When performing a **control test**: If the control bottle symbol and the flashing L1 or L2 DO NOT appear on the display with your control result, an error has occurred.

- Do not act on the control result.
- Discard the test strip and repeat the control test with a new test strip.

8

If you tested with Level 1 control solution, the L-1 screen appears. Press and hold the meter to confirm that you tested with Level 1 control solution.

or

If you tested with Level 2 control solution, the L-2 screen appears. Press and hold the meter button to confirm that you tested with Level 2 control solution.

Press the meter button to alternate between L-1 and L-2.

5_07945914061A_07945914061_01.indd 16 8/4/2021 11:05:37 AM

3 Control Tests

9

or

OK appears if the control result is within range.

Err appears if the control result is out of range.

10

Wipe the tip of the bottle with a tissue. Cap the bottle tightly.

Remove and discard the used test strip.

NOTE

The meter turns off 90 seconds after a successful test or 15 seconds after the test strip is removed, provided no other action is taken.

Understanding Out-of-Range Control Results

⚠ WARNING

Risk of a serious health incident
Incorrect test results can lead to wrong therapy recommendations.

The control ranges are printed on the test strip container label. If the control result is out of range, check that the meter and test strips are working properly. Check the following list to help solve the problem.

Troubleshooting Checks	Action
Were the test strips or control solutions expired?	Discard the test strips or control solution if either is past the use by date. If the control solution was opened more than 3 months ago, discard it. Repeat the control test with an unexpired test strip and an unexpired control solution.
Did you wipe the tip of the control solution bottle before use?	Wipe the tip of the bottle with a tissue. Repeat the control test with a new test strip and a fresh drop of control solution.

5_07945914061A_07945914061_01.indd 17 8/4/2021 11:05:39 AM

3 Control Tests

Troubleshooting Checks	Action
Were the caps on the test strip container and the control solution bottle always closed tightly?	Replace the test strips or control solution if you think either was uncapped for some time. Repeat the control test.
Was the test strip used immediately after it was removed from the test strip container?	Repeat the control test with a new test strip and a fresh drop of control solution.
Were the test strips and control solutions stored in a cool, dry place?	Repeat the control test with a properly stored test strip or control solution.
Did you follow the directions?	Read the chapter Control Tests and repeat the control test.
Did you choose the correct control solution level, either 1 or 2, when you performed the control test?	If you chose the wrong control solution level, you can still compare the control result to the range printed on the test strip container.
Are you still unsure of the problem?	Contact Roche.

5_07945914061A_07945914061_01.indd 18 8/4/2021 11:05:39 AM

4 Review Your Data

Overview

- The meter automatically stores at least 720 blood glucose results in memory, but only the last blood glucose result and your 7, 30 and 90-day averages can be viewed on the meter. To view stored blood glucose results, transfer them to a compatible software application.
- If more than 720 blood glucose tests are performed within a 90-day period, only the 720 most recent test results are included in the 90-day average.
- The meter automatically stores at least 30 control results in memory, but only the current control result can be viewed on the meter. To view stored control results, transfer them to a compatible software application.
- Control results are not viewable in memory and are not included in the 7, 30 or 90-day averages.

⚠ WARNING

Risk of a serious health incident
Using an individual test result in memory to change your therapy can lead to improper therapy decisions. Do not change your therapy based on an individual test result in memory. Talk to your healthcare professional before changing therapy based on test results in memory.

Reviewing Your Last Blood Glucose Result

1



With the meter off, press the meter button to turn the meter on. The last test result symbol ⮌ indicates that you are viewing your last blood glucose result.

Reviewing Your Test Result Averages

The meter uses all of your blood glucose results from the past 7, 30 or 90 days to calculate a single number. This number is called the average and helps you to understand your blood glucose results over a given period of time.

5_07945914061A_07945914061_01.indd 19 8/4/2021 11:05:41 AM

4 Review Your Data

1



With the meter off, press the meter button to turn the meter on. Press the meter button to view your 7, 30 and 90-day averages.

NOTE

If the memory contains a result that is HI or LO or is corrupt, then the time period and **day/ave** will flash to alert you that those results are not included in the average.

5_07945914061A_07945914061_01.indd 20 8/4/2021 11:05:47 AM

5 Connecting to a PC

Connecting the Meter to a PC Using a USB Cable

Follow the steps below to access diabetes management software on a PC. You can use this software to view stored results, set the meter time and date and change the target range.

1

Plug the small end of a USB cable into the meter.

2



Plug the large end of a USB cable into a USB port on the PC.

If the meter is off, it turns on.

3

Start the diabetes management software and initiate a data transfer.

4

The meter transfers the data to the software.

NOTE

The USB cable does not charge the meter battery. Remove the USB cable after you have finished.

5_07945914061A_07945914061_01.indd 21 8/4/2021 11:05:51 AM

Meter Maintenance

The meter automatically tests its own systems every time you turn it on and lets you know if something is wrong. See the Error Messages section of this chapter.

If you drop the meter or think the results are not accurate, contact Roche.

⚠ WARNING

Risk of life-threatening injuries
Keep new and used batteries away from children. See the warning in the Introduction of this User's Manual for additional information.

Changing the Battery

1



Open the child-resistant battery door by inserting a narrow object, such as a pen, into the slot (see image above). Push the tab in the direction of the arrow and lift the battery door up.

2



Remove the old battery. Press and hold the meter button for at least 2 seconds.

3



Slide the new battery under the tab with the **(+) side facing up**. Put the battery door back in place and snap it closed. Immediately discard the old battery.

NOTE

- Always have spare batteries.
- Battery life may vary due to factors such as temperature and battery manufacturer.
- The meter uses one 3-volt lithium battery, coin cell type CR2032. This type of battery can be found in many stores.
- Meter data is not lost when you replace the battery.

5_07945914061A_07945914061_01.indd 22 8/4/2021 11:05:56 AM

Checking the Meter Display

1



With the meter off, press the meter button to turn the meter on. Press the meter button twice to display the 30-day average screen.

2



Press and hold the meter button until all display segments appear. Check the meter display for missing segments. All segments should look like the picture above. If any segments are missing from the display, contact Roche, as there may be a problem with the meter.

Cleaning and Disinfecting the Meter

Keep the meter free of dust. If you need to clean or disinfect it, follow these guidelines carefully to help you get the best performance possible.

⚠ WARNING

Risk of infection
If the meter is being operated by a second person who is providing testing assistance to the user, the meter should be disinfected prior to use by the second person.

⚠ WARNING

Risk of a serious health incident

- Do not clean or disinfect the meter while performing a blood glucose or control test.
- Do not allow liquid to enter any openings in the meter.
- Do not spray anything directly onto the meter.
- Do not immerse the meter in liquid.

When to clean or disinfect the meter:

- Clean the meter to remove visible dirt or other material.
- Disinfect the meter between each patient use.

What to clean or disinfect:

- The area around slots and openings
- The meter display
- The entire meter surface

5_07945914061A_07945914061_01.indd 23 8/4/2021 11:06:00 AM

1

Make sure the meter is turned off.

Gently wipe the meter surface with a soft cloth slightly dampened (wring out any excess liquid) with one of these solutions:

To clean the meter

Mild dishwashing liquid mixed with water

To disinfect the meter

70 % isopropyl alcohol

Note: During disinfection, make sure that the meter surface stays wet with isopropyl alcohol for 2 minutes. Additional wiping of the surface with cloths dampened with isopropyl alcohol may be necessary.

2

Dry the meter thoroughly with a soft cloth.

Error Messages

⚠ WARNING

Risk of a serious health incident
Using an error message to change your therapy can lead to improper therapy decisions.

- Never make therapy decisions based on an error message.
- If you have any concerns or see any other error message, contact Roche.

NOTE

The error code and **Err** will alternate on the display for all coded error messages (E-1 through E 14).



The meter will not turn on or the display is blank.

- Battery is dead.

Insert new battery.

- Display is damaged. / Meter is defective.

Contact Roche.

- Extreme temperatures.

Move the meter to a location with a more moderate temperature.

The meter is connected to a PC and a blood glucose or control test cannot be performed.

EITHER remove the USB cable and perform a blood glucose or control test OR remove the test strip and start a data transfer.

5_07945914061A_07945914061_01.indd 24 8/4/2021 11:06:01 AM

The battery is almost out of power.

Change the battery now. If the symbol reappears after the battery has been replaced, remove the battery again, press and hold the meter button for at least 2 seconds, then reinsert the battery.

(The error code and **Err** will alternate on the display.)

The test strip may be damaged, not properly inserted or was previously used.

Remove and reinsert the test strip or replace it if damaged or previously used.

(The error code and **Err** will alternate on the display.)

A meter or test strip error has occurred.

This error message could appear if the cap on the test strip container was not closed tightly. The test strips may have been damaged due to improper storage or handling.

Never make therapy decisions based on an error message.

Repeat the blood glucose test. If a second E-3 error message appears, perform a control test with the control solution and a new test strip. See the Performing a Control Test section in the chapter Control Tests. If you continue to receive an E-3 error message, use an alternate method for testing your blood glucose, such as a back-up meter and test strip. If the alternate method gives an extremely high blood glucose result or if an alternate method is not available, **contact your healthcare professional immediately**.

In rare cases, an E-3 error message may indicate that your blood glucose is extremely high and above the system's measuring range. **See the Unusual Blood Glucose Results section in the chapter Blood Glucose Tests for other possible causes of the error message.**

5_07945914061A_07945914061_01.indd 25 8/4/2021 11:06:03 AM

(The error code and **Err** will alternate on the display.)

Not enough blood or control solution was drawn into the test strip for measurement or was applied after the test had started.

Discard the test strip and repeat the blood glucose or control test.

(The error code and **Err** will alternate on the display.)

Blood or control solution was applied to the test strip before the flashing drop symbol appeared on the display.

Discard the test strip and repeat the blood glucose or control test.

(The error code and **Err** will alternate on the display.)

An electronic error occurred.

Remove the battery, press and hold the meter button for at least 2 seconds and reinsert the battery. Perform a blood glucose or control test.

(The error code and **Err** will alternate on the display.)

The temperature is above or below the proper range for the system.

Refer to the test strip package insert for system operating conditions. Move to an area with the appropriate conditions and repeat the blood glucose or control test. Do not artificially heat or cool the meter.

(The error code and **Err** will alternate on the display.)

The battery may be out of power.

Turn the meter back on. **If you are in a cold environment, move to a location with a more moderate temperature and retest.** If the message continues to appear after several attempts, replace the battery. If the message reappears after the battery has been replaced, remove the battery, press and hold the meter button for at least 2 seconds, then reinsert the battery.

5_07945914061A_07945914061_01.indd 26 8/4/2021 11:06:07 AM

(The error code and **Err** will alternate on the display.)

The meter's internal clock may need to be reset.

Press the meter button. If the time and date were set, you will need to reset them. See the Time and Date section in the chapter Your New System.

(The error code and **Err** will alternate on the display.)

The test strip may be damaged.

Repeat the blood glucose or control test with a new test strip.

(The error code and **Err** will alternate on the display.)

Your blood sample may contain a high level of ascorbate.

Contact your healthcare professional.

(The error code and **Err** will alternate on the display.)

Fluid or foreign material may be present in the test strip slot.

Remove and reinsert the test strip or repeat the blood glucose or control test with a new test strip. If the error persists, contact Roche.

(The error code and **Err** will alternate on the display.)

An electronic error has occurred.

Contact Roche.

5_07945914061A_07945914061_01.indd 27 8/4/2021 11:06:11 AM

Blood glucose may be higher than the measuring range of the system.

See the Unusual Blood Glucose Results section in the chapter Blood Glucose Tests.

Blood glucose may be lower than the measuring range of the system.

See the Unusual Blood Glucose Results section in the chapter Blood Glucose Tests.

5_07945914061A_07945914061_01.indd 28 8/4/2021 11:06:13 AM

7 Technical Information

Product Limitations

See the literature packaged with the test strips and control solutions for the latest information on product specifications and limitations.

Specifications	
Blood volume Sample type Measuring time Measuring range Test strip storage conditions System operating conditions	Refer to the test strip package insert.
Meter storage conditions	Temperature: -25–70 °C
Memory capacity	1 blood glucose result and 7, 30 and 90-day test result averages (at least 720 blood glucose results and at least 30 control results are also stored and can be viewed with external software).
Automatic off	90 seconds after performing a test, 15 seconds after a test strip is removed or 5 seconds from the last test result screen.
Power supply	One 3-volt lithium battery (coin cell type CR2032)
Display	LCD
Dimensions	77.1 × 48.6 × 15.3 mm (LWH)
Weight	Approx. 40 g (with battery)
Construction	Hand-held
Protection class	III
Meter type	The Accu-Chek Instant S meter is suitable for continuous operation.
Control solution storage conditions	Refer to the control solution package insert.
Continua CERTIFIED Interfaces	USB: micro-B connector; Continua Certified® to a Continua Certified manager.

5_07945914061A_07945914061_01.indd 29 8/4/2021 11:06:13 AM

Electromagnetic Compatibility – The meter meets the electromagnetic emission requirements as per EN 61326-2-6. Its electromagnetic emission is thus low. Interference on other electrically-driven equipment is not anticipated.

Performance Analysis – Refer to the test strip package insert.

Test Principle – Refer to the test strip package insert.

Communication Protocol – The Accu-Chek Instant S blood glucose meter is Continua Certified. Continua Certified signifies that this product complies with applicable IEEE 11073-10417 standards and that it has been tested and certified against the 2015 Continua Design Guidelines.

Product Safety Information

NOTE

- Strong electromagnetic fields may interfere with the proper operation of the meter. Do not use the meter close to sources of strong electromagnetic radiation.
- To avoid electrostatic discharge, do not use the meter in a very dry environment, especially one in which synthetic materials are present.

Discarding the Meter

⚠ WARNING

Risk of infection
During blood glucose testing, the meter itself may come into contact with blood. Used meters therefore carry a risk of infection.

Keep new and used batteries away from children. See the warning in the Introduction of this User's Manual for additional information

Before discarding the meter, remove the battery or batteries.

Discard used meters according to the regulations applicable in your country. Contact the local council and authority for information about correct disposal.

The meter falls outside the scope of the European Directive 2012/19/EU (Directive on waste electrical and electronic equipment (WEEE)).

Discard used batteries according to local environmental regulations.

This product includes an internal constituent containing a Substance of Very High Concern (SVHC), Hexabromocyclododecane (CAS 25637-99-4), in a concentration above 0.1 % weight by weight, as identified under REACH and added to the Candidate List.

There is no direct exposure to the substance and therefore no risk when the instrument is operated according to the instructions for use.

5_07945914061A_07945914061_01.indd 30 8/4/2021 11:06:13 AM

7 Technical Information

Explanation of Symbols

These symbols may appear on the packaging, on the type plate and in the instructions for the Accu-Chek Instant S meter.

	Consult instructions for use or consult electronic instructions for use
	Caution, refer to safety-related notes in the instructions for use accompanying this product.
	Temperature limit
	Use by
	Keep new and used batteries away from children.
	3-volt coin cell type CR2032
	Biological risks – used meters carry a risk of infection.
	Date of manufacture
IVD	In vitro diagnostic medical device
GTIN	Global Trade Item Number
	Device for self-testing
	Device for near-patient testing
	Manufacturer
UDI	Unique device identifier
REF	Catalogue number
SN	Serial number

5_07945914061A_07945914061_01.indd 31 8/4/2021 11:06:20 AM

LOT	Batch code
CE	Complies with the provisions of the applicable EU Legislation

Additional Supplies

Test Strips: Accu-Chek Instant test strips

Control Solutions: Accu-Chek Instant control solutions

Information for Healthcare Professionals

This system can be used in professional healthcare environments such as doctors' offices, general wards, in suspected cases of diabetes and in emergency cases.

⚠ WARNING

Risk of infection
Any object coming into contact with human blood is a potential source of infection. Healthcare professionals should follow the infection control procedures appropriate for your facility. Refer to the test strip package insert for additional healthcare professional information.

Sample Handling

Always wear gloves when handling blood-contaminated items. Always adhere to the recognised procedures for handling objects that are potentially contaminated with human material. Follow the hygiene and safety policy of your laboratory or institution. Prepare the selected blood collection site per facility policy.

Refer to the test strip package insert for additional information regarding acceptable sample types, anticoagulants and handling instructions.

Recommending Alternative Site Testing to Patients

Decisions about whether to recommend Alternative Site Testing (AST) should take into account the motivation and knowledge level of the patient and his or her ability to understand the considerations relative to diabetes and AST. If you are considering recommending AST for your patients, you need to understand that there is a potential for a significant difference between fingertip or palm test results and test results obtained from the forearm or upper arm. The difference in capillary bed concentration and blood perfusion throughout the body can lead to sample site-to-site differences in blood glucose results. These physiological effects vary between individuals and can vary within a single individual based upon his or her behavior and relative physical condition.

Our studies involving alternative site testing of adults with diabetes show that most persons will find their glucose level changes more quickly in blood from the

5_07945914061A_07945914061_01.indd 32 8/4/2021 11:06:20 AM

fingertip or palm than in blood from the forearm or upper arm. This is especially important when blood glucose levels are falling or rising rapidly. If your patient is used to making therapy decisions based upon fingertip or palm test results, he or she should consider the delay or lag time, affecting the test results obtained with blood from the forearm or upper arm.

5_07945914061A_07945914061_01.indd 33 8/4/2021 11:06:20 AM

8 Guarantee

Guarantee

The statutory provisions on rights in consumer goods sales in the country of purchase shall apply.

5_07945914061A_07945914061_01.indd 34 8/4/2021 11:06:20 AM

Index

A
alternative site testing 12, 32
averages 8, 19

B
battery, changing 22
battery type 22, 29
blood glucose results, unusual 13
blood glucose test, performing 10
button, functions 5

C
control results, out-of-range 17
control solution 15
control test, performing 15

D
data transfer, USB 21

E
error messages 24

G
guarantee 34

H
healthcare professionals 32
high blood glucose 14
hyperglycaemia 14
hypoglycaemia 14

L
last test result 8
low blood glucose 14

M
maintenance, meter 22
meter, cleaning and disinfecting 23

P
product limitations 29
product safety information 30
product specifications 29

S
supplies 6, 32
symbols
    IVD 31
    meter display 8

T
target range 11
target range indicator 5, 11
technical information 29
test strips 9
time and date 7

U
use by date 10, 15

5_07945914061A_07945914061_01.indd 35 8/4/2021 11:06:20 AM

LAST UPDATE: 2021-06

Australia
Roche Diabetes Care Australia Pty. Limited
Accu-Chek Enquiry Line: 1800 251 816
www.accu-chek.com.au

India
Accu-Chek Customer Care Helpline: 1800 120 6020 (Toll-free)
(Monday to Saturday, 7:30 AM to 9:30 PM)
SMS 'HELP' to 5757535*
*Charges applicable as per your network provider
Email: india.accu-chek@roche.com
www.accu-chek.in

Malaysia
Customer Support and Service Centre:
Toll-Free Line: 1-800-88-1313
www.accu-chek.com.my

Roche Pakistan Limited
37-B, 1st Floor, Block 6,
PECHS Karachi - 75400 Pakistan
Contact: 0800-76243
www.accu-chek.com.pk

Philippines
Accu-Chek Customer Hotline:
Phone: (02) 7718 7575 / (02) 7718 7588 /
+63 917 8978000
www.accu-chek.com.ph

Roche Diabetes Care South Africa (Pty) Ltd.
Hertford Office Park, 90 Bekker Road
Vorna Valley, 1686
South Africa
Tel: +27 (11) 504 4600
Email: info@accu-chek.co.za
www.accu-chek.co.za

United Arab Emirates
Roche Diabetes Care Middle East FZCO
Tel: +971 (0) 4 816 9165
www.rochediabetescareme.com
or contact Roche local authorized representative in your country

대한민국
고객 지원 및 서비스 센터
무료상담전화: 080-909-2222
(월-금 오전 9시 ~ 오후 5시)
www.accu-chek.co.kr

Roche Diabetes Care GmbH
Sandhofer Strasse 116
68305 Mannheim, Germany
www.accu-chek.com

IVD CE 0123

ACCU-CHEK and ACCU-CHEK INSTANT are trademarks of Roche.

Continua CERTIFIED

CONTINUA, the CONTINUA logos and CONTINUA CERTIFIED are trademarks, service marks or certification marks of the Continua Health Alliance. CONTINUA is a registered trademark in some, but not all countries in which this product is distributed.

All other product names and trademarks are the property of their respective owners.

© 2021 Roche Diabetes Care

07945914061A-0821

07945914061(01)-0821